



**BID BULLETIN NO. 01**

Date: **23 June 2023**

ITB No.: **bac-23-0614B**

Project Name: **Supply and Delivery of One (1) Unit Compact Mobile C-Arm –PCCH**

**ABC: Php10,000,000.00**

**To all prospective bidders:**

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 14 June 2023. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.

Item	Previous Specification/ Clarification/Request to Consider	Amendment/Response to Clarification								
<b>1</b>	<p style="text-align: center;"><b>PREVIOUS SECTION VII TECHNICAL SPECIFICATIONS</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">ITEM</th> <th style="width: 85%;">SPECIFICATION</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>           xxxxxxxxxxxxxxxx             X-ray Generator: Minimum Specification or better &gt; Nominal peak output power should be at least 2.3W             xxxxxxxxxxxxxxxx             Hand switch should conform to at least IPX 3 standard and should be protected against dust and splashing of water.             xxxxxxxxxxxxxxxx         </td> </tr> </tbody> </table>	ITEM	SPECIFICATION	1	xxxxxxxxxxxxxxxx  X-ray Generator: Minimum Specification or better > Nominal peak output power should be at least 2.3W  xxxxxxxxxxxxxxxx  Hand switch should conform to at least IPX 3 standard and should be protected against dust and splashing of water.  xxxxxxxxxxxxxxxx	<p style="text-align: center;"><b>AMENDMENT SECTION VII TECHNICAL SPECIFICATIONS</b></p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">ITEM</th> <th style="width: 85%;">SPECIFICATION</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>           xxxxxxxxxxxxxxxx             X-ray Generator: Minimum Specification or better &gt; Nominal peak output power should be at least <b>2.3kW</b>             xxxxxxxxxxxxxxxx             Hand switch should conform to at least <b>IPX 1</b> standard and should be protected against dust and splashing of water.             xxxxxxxxxxxxxxxx         </td> </tr> </tbody> </table>	ITEM	SPECIFICATION	1	xxxxxxxxxxxxxxxx  X-ray Generator: Minimum Specification or better > Nominal peak output power should be at least <b>2.3kW</b>  xxxxxxxxxxxxxxxx  Hand switch should conform to at least <b>IPX 1</b> standard and should be protected against dust and splashing of water.  xxxxxxxxxxxxxxxx
ITEM	SPECIFICATION									
1	xxxxxxxxxxxxxxxx  X-ray Generator: Minimum Specification or better > Nominal peak output power should be at least 2.3W  xxxxxxxxxxxxxxxx  Hand switch should conform to at least IPX 3 standard and should be protected against dust and splashing of water.  xxxxxxxxxxxxxxxx									
ITEM	SPECIFICATION									
1	xxxxxxxxxxxxxxxx  X-ray Generator: Minimum Specification or better > Nominal peak output power should be at least <b>2.3kW</b>  xxxxxxxxxxxxxxxx  Hand switch should conform to at least <b>IPX 1</b> standard and should be protected against dust and splashing of water.  xxxxxxxxxxxxxxxx									



	Should have a focal spot nominal value of at least 0.6 and 1  XXXXXXXXXXXXXXXXXXXX	Should have a focal spot nominal value of at least <b>0.5 to 0.6</b> and 1  XXXXXXXXXXXXXXXXXXXX
<b>2</b>	<b>PREVIOUS TERMS OF REFERENCE</b>  -Conduct hands-on actual demonstration of the model/unit being offered at Pasig City Children’s Hospital during Post-evaluation after bid opening	<b>AMENDMENT TERMS OF REFERENCE</b>  - Conduct hands-on actual demonstration of the model/unit being offered <b>at Pasig City Children’s Hospital</b> during <b>Post-qualification</b>

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as “Modified Bid”) and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

Conforme:

  
 \_\_\_\_\_  
 END-USER REPRESENTATIVE  
 Signature Over Printed Name

  
**ATTY. JOSEPHINE C. LATI-BAGAOISAN**  
 Chairperson *dm*