



BID BULLETIN NO. 01

Date: **09 August 2023**

ITB No.: **bac-23-0726a**

Project Name: **Supply and Delivery of Various Laparoscopic Instrument
- PCGH**

ABC: Php1,328,239.00

To all prospective bidders:

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 26 July 2023. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.


Item	Previous Specification/ Clarification/Request to Consider	Amendment/Response to Clarification																						
1	PREVIOUS SECTION VI. SCHEDULE OF REQUIREMENTS, SECTION VII. TECHNICAL SPECIFICATIONS AND ITEMS TO BE BID	AMENDMENT TO SECTION VI. SCHEDULE OF REQUIREMENTS, SECTION VII. TECHNICAL SPECIFICATIONS AND ITEMS TO BE BID																						
	<table border="1"><thead><tr><th>ITEM NO.</th><th>SPECIFICATION</th></tr></thead><tbody><tr><td>10</td><td>De Bakey 5mm, jaw length 40mm, insert</td></tr><tr><td>12</td><td>Johan grasping forceps 5mm, 22mm, slightly flat jaws, insert</td></tr><tr><td>15</td><td>5 mm sheath, uses special radel insulation that minimizes the risk of arcing</td></tr><tr><td>18</td><td>5mm sheath, uses special radel insulation that minimizes the risk</td></tr><tr><td>21</td><td>5mm sheath, uses special radel insulation that minimizes the risk of arcing</td></tr></tbody></table>	ITEM NO.	SPECIFICATION	10	De Bakey 5mm, jaw length 40mm, insert	12	Johan grasping forceps 5mm, 22mm, slightly flat jaws, insert	15	5 mm sheath, uses special radel insulation that minimizes the risk of arcing	18	5mm sheath, uses special radel insulation that minimizes the risk	21	5mm sheath, uses special radel insulation that minimizes the risk of arcing	<table border="1"><thead><tr><th>ITEM NO.</th><th>SPECIFICATION</th></tr></thead><tbody><tr><td>10</td><td>De Bakey 5mm, jaw length 40-45mm, insert</td></tr><tr><td>12</td><td>Johan grasping forceps 5mm, 22-25mm, slightly flat jaws, insert</td></tr><tr><td>15</td><td>5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing</td></tr><tr><td>18</td><td>5mm sheath, uses special radel insulation that minimizes the risk or its equivalent risk of arcing</td></tr></tbody></table>	ITEM NO.	SPECIFICATION	10	De Bakey 5mm, jaw length 40-45mm , insert	12	Johan grasping forceps 5mm, 22-25mm , slightly flat jaws, insert	15	5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing	18	5mm sheath, uses special radel insulation that minimizes the risk or its equivalent risk of arcing
ITEM NO.	SPECIFICATION																							
10	De Bakey 5mm, jaw length 40mm, insert																							
12	Johan grasping forceps 5mm, 22mm, slightly flat jaws, insert																							
15	5 mm sheath, uses special radel insulation that minimizes the risk of arcing																							
18	5mm sheath, uses special radel insulation that minimizes the risk																							
21	5mm sheath, uses special radel insulation that minimizes the risk of arcing																							
ITEM NO.	SPECIFICATION																							
10	De Bakey 5mm, jaw length 40-45mm , insert																							
12	Johan grasping forceps 5mm, 22-25mm , slightly flat jaws, insert																							
15	5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing																							
18	5mm sheath, uses special radel insulation that minimizes the risk or its equivalent risk of arcing																							

24	5 mm sheath, uses special radel insulation that minimizes the risk of arcing	21	5mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing
27	5 mm sheath, uses special radel insulation that minimizes the risk of arcing	24	5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing
30	5 mm sheath, uses special radel insulation that minimizes the risk of arcing	27	5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing
33	5 mm sheath, uses special radel insulation that minimizes the risk of arcing	30	5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing
36	5 mm sheath, uses special radel insulation that minimizes the risk of arcing	33	5 mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing
		36	10mm sheath, uses special radel insulation that minimizes the risk of arcing or its equivalent risk of arcing

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as "Modified Bid") and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

Conforme:



 END-USER REPRESENTATIVE
 Signature Over Printed Name


ATTY. JOSEPHINE C. LATI-BAGAOISAN
 Chairperson