



BID BULLETIN NO. 02

Date: **09 October 2023**

ITB No.: **bac-23-0920A**

Project Name: **Supply and Delivery of Two (2) Units Ventilator Machine
– PCGH Nursing Service Office**

ABC: Php3,400,000.00

To all prospective bidders:

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 20 September 2023. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.

Item	Previous Specification/ Clarification/Request to Consider	Amendment/Response to Clarification
1	<p>PREVIOUS SECTION I. INVITATION TO BID</p> <p>xxx xxx xxx</p> <p>7. Bids must be duly received by the Procurement Management Office through manual submission at the office address indicated below, on or before <i>October 11, 2023, 9:30 A.M.</i> Late bids shall not be accepted.</p> <p>xxx xxx xxx</p> <p>9. Bid opening shall be on <i>October 11, 2023, 10:00 A.M</i> at the given address below. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>	<p>AMENDMENT TO SECTION I. INVITATION TO BID</p> <p>xxx xxx xxx</p> <p>7. Bids must be duly received by the Procurement Management Office through manual submission at the office address indicated below, on or before October 16, 2023, 9:30 A.M. Late bids shall not be accepted.</p> <p>xxx xxx xxx</p> <p>9. Bid opening shall be on October 16, 2023, 10:00 A.M at the given address below. Bids will be opened in the presence of the bidders'</p>

	<p style="text-align: center;">xxx xxx xxx</p>	<p>representatives who choose to attend the activity.</p> <p style="text-align: center;">xxx xxx xxx</p>
<p style="text-align: center;">2</p>	<p>PREVIOUS REMINDERS</p> <p style="text-align: center;">xxx xxx xxx</p> <p>BIDDING DOCUMENTS AND AVAILABILITY FEE</p> <ul style="list-style-type: none"> • Bidding Documents: <ul style="list-style-type: none"> ➤ September 20, 2023 to October 11, 2023 until 9:30 AM ➤ 8:00 am to 5:00 pm and upon payment of applicable fees for the Bidding Documents at the City Treasurer’s Office <p style="text-align: center;">xxx xxx xxx</p> <p>REMINDERS:</p> <ul style="list-style-type: none"> • The deadline for the submission of bid is on October 11, 2023 (Wednesday) at 9:30 AM at the BAC Secretariat’s Office, 4th Floor Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. The digital clock at the BAC Secretariat’s Office that is set to the Philippine Time (PhST) shall be used as reference in determining the time for the submission of bids, hence participating bidders are advised to synchronize their timepiece with the said digital clock. Late bids or those who submitted after 9:30 AM of October 11, 2023 (Wednesday) shall not be accepted. 	<p>AMENDMENT TO REMINDERS</p> <p style="text-align: center;">xxx xxx xxx</p> <p>BIDDING DOCUMENTS AND AVAILABILITY FEE</p> <ul style="list-style-type: none"> • Bidding Documents: <ul style="list-style-type: none"> ➤ September 20, 2023 to October 16, 2023 until 9:30 AM ➤ 8:00 am to 5:00 pm and upon payment of applicable fees for the Bidding Documents at the City Treasurer’s Office <p style="text-align: center;">xxx xxx xxx</p> <p>REMINDERS:</p> <ul style="list-style-type: none"> • The deadline for the submission of bid is on October 16, 2023 (Monday) at 9:30 AM at the BAC Secretariat’s Office, 4th Floor Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. The digital clock at the BAC Secretariat’s Office that is set to the Philippine Time (PhST) shall be used as reference in determining the time for the submission of bids, hence participating bidders are advised to synchronize their timepiece with the said digital clock. Late bids or those who submitted after 9:30 AM of October 16, 2023 (Monday) shall not be accepted.


	XXX XXX XXX		XXX XXX XXX											
	<ul style="list-style-type: none"> Bid opening shall be on October 11, 2023 (Wednesday) at 10:00 AM at 7th Floor Meeting Room, Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. Bids will be opened in the presence of the bidders' representatives who choose to attend. 		<ul style="list-style-type: none"> Bid opening shall be on October 16, 2023 (Monday) at 10:00 AM at 7th Floor Meeting Room, Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. Bids will be opened in the presence of the bidders' representatives who choose to attend. 											
	XXX XXX XXX		XXX XXX XXX											
3	PREVIOUS SECTION VII. TECHNICAL SPECIFICATIONS	AMENDMENT TO VII. TECHNICAL SPECIFICATIONS												
	<table border="1"> <thead> <tr> <th style="text-align: center;">ITEM</th> <th style="text-align: center;">SPECIFICATION</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td style="text-align: center;">xxxxxxx Dimensions and weight: "Main unit At least - 34 cm x 26 cm x14 cm Weight. At least 3Kg - 10 kg xxxxxxx</td> </tr> <tr> <td style="text-align: center;">1</td> <td style="text-align: center;">xxxxxxx " Adjustable parameters "Tidal Volume ((Vt). Neonate to Adult at least with :MI 2 to 4000 Adult mode "Minute ventilation (MV), I/min 0.02-80 "Oxygen concentration (FiO2), % 21-100 "LPO target O2 concentration, % 21-100 "Breathe rate: Neonate to adult 1-150 mode "Inspiratory time about 0.1-48 "Inspiratory pause about 0-70% of inspiratory time "Sigh 1:50, 1:100 "PEEP Sigh, cmH20 0-15 every minutes</td> </tr> </tbody> </table>	ITEM	SPECIFICATION	1	xxxxxxx Dimensions and weight: "Main unit At least - 34 cm x 26 cm x14 cm Weight. At least 3Kg - 10 kg xxxxxxx	1	xxxxxxx " Adjustable parameters "Tidal Volume ((Vt). Neonate to Adult at least with :MI 2 to 4000 Adult mode "Minute ventilation (MV), I/min 0.02-80 "Oxygen concentration (FiO2), % 21-100 "LPO target O2 concentration, % 21-100 "Breathe rate: Neonate to adult 1-150 mode "Inspiratory time about 0.1-48 "Inspiratory pause about 0-70% of inspiratory time "Sigh 1:50, 1:100 "PEEP Sigh, cmH20 0-15 every minutes	<table border="1"> <thead> <tr> <th style="text-align: center;">ITEM</th> <th style="text-align: center;">SPECIFICATION</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td style="text-align: center;">xxxxxxx Dimensions and weight: "Main unit - Manufacturer's standard Weight Manufacturer's standard xxxxxxx</td> </tr> <tr> <td style="text-align: center;">1</td> <td style="text-align: center;">xxxxxxx " Adjustable parameters "Tidal Volume ((Vt). Neonate to Adult at least with : MI 2 to 2000 Adult mode "Minute ventilation (MV), I/min 0.02-80 "Oxygen concentration (FiO2), % 21-100 "LPO target O2 concentration, % 21-100 "Breathe rate: Neonate to adult 1-150 mode "Inspiratory time about 0.1-10s "Inspiratory pause about 0-60% of inspiratory time "Sigh 1:50, 1:100 "PEEP Sigh, cmH20 0-15 every minutes</td> </tr> </tbody> </table>	ITEM	SPECIFICATION	1	xxxxxxx Dimensions and weight: "Main unit - Manufacturer's standard Weight Manufacturer's standard xxxxxxx	1	xxxxxxx " Adjustable parameters "Tidal Volume ((Vt). Neonate to Adult at least with : MI 2 to 2000 Adult mode "Minute ventilation (MV), I/min 0.02-80 "Oxygen concentration (FiO2), % 21-100 "LPO target O2 concentration, % 21-100 "Breathe rate: Neonate to adult 1-150 mode "Inspiratory time about 0.1- 10s "Inspiratory pause about 0- 60% of inspiratory time "Sigh 1:50, 1:100 "PEEP Sigh, cmH20 0-15 every minutes
ITEM	SPECIFICATION													
1	xxxxxxx Dimensions and weight: "Main unit At least - 34 cm x 26 cm x14 cm Weight. At least 3Kg - 10 kg xxxxxxx													
1	xxxxxxx " Adjustable parameters "Tidal Volume ((Vt). Neonate to Adult at least with :MI 2 to 4000 Adult mode "Minute ventilation (MV), I/min 0.02-80 "Oxygen concentration (FiO2), % 21-100 "LPO target O2 concentration, % 21-100 "Breathe rate: Neonate to adult 1-150 mode "Inspiratory time about 0.1-48 "Inspiratory pause about 0-70% of inspiratory time "Sigh 1:50, 1:100 "PEEP Sigh, cmH20 0-15 every minutes													
ITEM	SPECIFICATION													
1	xxxxxxx Dimensions and weight: "Main unit - Manufacturer's standard Weight Manufacturer's standard xxxxxxx													
1	xxxxxxx " Adjustable parameters "Tidal Volume ((Vt). Neonate to Adult at least with : MI 2 to 2000 Adult mode "Minute ventilation (MV), I/min 0.02-80 "Oxygen concentration (FiO2), % 21-100 "LPO target O2 concentration, % 21-100 "Breathe rate: Neonate to adult 1-150 mode "Inspiratory time about 0.1- 10s "Inspiratory pause about 0- 60% of inspiratory time "Sigh 1:50, 1:100 "PEEP Sigh, cmH20 0-15 every minutes													

	"Inspiratory pressure (P control) / 0-100 Adult mode "Pressure support (PS), cmH2O 0-80 Pediatric mode "0-60 Neonatal mode "I:E ratio (I:E) 1:10-4:1 "PEEP/CPAP, cm H2O 0-50 xxx xxx xxx	"Inspiratory pressure (P control) / 0-100 Adult mode "Pressure support (PS), cmH2O 0-80 Pediatric mode "0-60 Neonatal mode "I:E ratio (I:E) 1:10-4:1 "PEEP/CPAP, cm H2O 0-45 adjustable parameters xxx xxx xxx

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as "Modified Bid") and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

Conforme:



 END-USER REPRESENTATIVE
 Signature Over Printed Name

F: 
ATTY. JOSEPHINE C. LATI-BAGAOISAN 
 Chairperson