



Bid Notice Abstract

Request for Quotation (RFQ)

Reference Number 10328647
Procuring Entity CITY OF PASIG
Title Supply and Delivery of 1 unit of Cautery Machine – City Health Office
Area of Delivery Metro Manila

| | | |
|--|------------------------------|---------------------|
| Solicitation Number: R2 100-23-05-1180 | Status | Active |
| Trade Agreement: Implementing Rules and Regulations | | |
| Procurement Mode: Negotiated Procurement - Small Value Procurement (Sec. 53.9) | Associated Components | 1 |
| Classification: Goods | | |
| Category: Hospital / Medical Equipment | Bid Supplements | 1 |
| Approved Budget for the Contract: PHP 550,000.00 | | |
| Delivery Period: 14 Day/s | Document Request List | 11 |
| Client Agency: | | |
| Contact Person: ATTY. PONCE MIGUEL D. LOPEZ Officer In Charge, Procurement Management Office Pasig City Hall, Caruncho Avenue, Barangay San Nicolas, Pasig City Metro Manila Philippines 1600 63-2-86431111 Ext.1461 bidsandawards@pasigcity.gov.ph | Date Published | 20/11/2023 |
| | Last Updated / Time | 28/11/2023 10:13 AM |
| | Closing Date / Time | 01/12/2023 09:00 AM |
| | | |

Description

Items Quantity / Units

- 1 Cautery Machine,
- With isolated output electrosurgical energy system
- With at least 15.7cm (6.2 inch) LCD touchscreen monitor, must be at least in soft touch function
- should be sturdy and can be placed on a table top or with sturdy movable stand
- must have for the following settings: Cut, Coag, Bipolar and shared coag.
- must have recall program from the last shutdown for power settings.
- must have auto-bipolar function and meter reading
- must have neutral electrode alarm that is activated in case contact with patient is broken for maximum safety.
- must have at least one (1) bipolar output plug and at least two (2) monopolar plug
- must have technology that insulates the system from interference from other electrical equipment
- return pads must be compatible with all return pads currently available in the philippine market
- must be supplied with foot pedals: one (1) for monopolar (CUT and COAG) and one (1) for independent
- POWER SUPPLY: Auto-Voltage or 220V, 60Hz

