



Bid Notice Abstract

Request for Quotation (RFQ)

Reference Number 10436285
Procuring Entity CITY OF PASIG
Title Supply and Delivery of Various Medical Equipment (Item No. 5,14,15) – Pasig City Children’s Hospital
Area of Delivery Metro Manila

Solicitation Number: R1 100-23-05-1961	Status	Active
Trade Agreement: Implementing Rules and Regulations		
Procurement Mode: Negotiated Procurement - Small Value Procurement (Sec. 53.9)	Associated Components	1
Classification: Goods		
Category: Medical Supplies and Laboratory Instrument	Bid Supplements	1
Approved Budget for the Contract: PHP 576,000.00		
Delivery Period: 90 Day/s	Document Request List	11
Client Agency:		
Contact Person: ATTY. PONCE MIGUEL D. LOPEZ Officer In Charge, Procurement Management Office Pasig City Hall, Caruncho Avenue, Barangay San Nicolas, Pasig City Metro Manila Philippines 1600 63-2-86431111 Ext.1461 bidsandawards@pasigcity.gov.ph	Date Published	20/12/2023
	Last Updated / Time	05/01/2024 14:12 PM
	Closing Date / Time	10/01/2024 09:00 AM

Description

Items Quantity / Units

- 5 Diagnostic Set
OTOSCOPE:
- Sealed Pneumatic otoscopy: Yes
- Illumination type: ring fiber optics
- Lamp life: at least 50 yrs
- Magnification: ~ 5X
- Total light output: at least 5200 lux; 2.5 times brighter than traditional halogen scopes.
- Color rendering index (CRI): 95 typical; R9: 90 typical; Color Temperature: 3400K - 4500K 7 years
- Dimensions handle at least: 14 cm x 3 cm x 3 cm
- Weight at least 140 grams
- With Rechargeable Battery
- INCLUSION: (1) hard case, (1) charger
- OPHTHALMOSCOPE SET :

With halogen light

- Aperture: micro, small, large, fixation target, slit, red-free filter
- At least 28 focusing lenses with ranges of -25 to +40 diopters

9 set

14 Pouch Sealer with Cutter

- compatible with sterilization rolls for manual cutting and sealing.
- Can meet the needs of high temperature steam sterilization, low temperature ethylene oxide, hydrogen peroxide plasma and radiation sterilization.

- Compatible with high density polyethylene material.

- Working temperature: 60-220°C adjustable

- Precision of Temperature: $< \pm 2 \% \text{ } ^\circ\text{C}$

INCLUSION: (1pc) Power Supply with universal socket adapter 2 unit

15 Heavy Duty Pouch Sealer (CONVEYOR TYPE)

- Machine size: At least 850mm x 420mm x 360mm

- Machine Weight: At least 19 kg but not more than 25kg

- Maximum loading: 3kg

- Seal all types of pouches using continuous sealer with rolling tray

- Automatic operation.

- Can set temperature and speed.

- Working speed: 0-12m/min.

- Sealing Width: 12mm

- Temperature Range: 0-300°C

- Working speed: 0-12m/min.

- Power: Not more than 500w 1 unit

PRICE QUOTATION/S SHOULD BE PRINTED ON COMPANY'S OFFICIAL LETTERHEAD TOGETHER WITH THE FOLLOWING UPDATED DOCUMENTS;

-Mayor's/Business Permit

-PhilGEPS Registration Number

-Income/Business Tax Return

-Accomplished and notarized Omnibus Sworn Statement

([https://www.gppb.gov.ph/assets/forms/Omnibus%20Sworn%20Statement \(Revised\).docx](https://www.gppb.gov.ph/assets/forms/Omnibus%20Sworn%20Statement%20(Revised).docx))

-Proof of Authorization: Secretary's Certificate if corporation, or Special Power Of Attorney, if individual

NOTE:

TO BE SUBMITTED SEALED AND LABELED ON A LONG BROWN ENVELOPE, FOLLOWING THIS FORMAT AS FOLLOWS:

FOR:

ATTY. JOSEPHINE C. LATI-BAGAOISAN

BAC Chairperson

THRU:

ATTY. PONCE MIGUEL D. LOPEZ

Officer In Charge, Procurement Management Office

4th Floor, Pasig City Hall,

Caruncho Avenue,

Pasig City

DATE : _____

COMPANY'S NAME : _____

PhilGEPS REFERENCE NUMBER : _____

PROJECT TITLE : _____

Remarks

New closing date, January 10, 2024 at 9:00 AM

Please be guided accordingly

Created by ATTY. PONCE MIGUEL D. LOPEZ

Date Created 19/12/2023

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.

