



BID BULLETIN NO. 01

Date: **08 March 2024**

ITB No.: **bac-24-0219g**

Project Name: **Supply and Delivery of Various Medical Equipment - PCCH (Rebid)**

ABC: Php3,300,000.00

To all prospective bidders:

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 19 February 2024. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.

Item	Previous Specification/ Clarification/Request to Consider	Amendment/Response to Clarification
1	<p>PREVIOUS SECTION I. INVITATION TO BID</p> <p>xxx xxx xxx</p> <p>7. Bids must be duly received by the Procurement Management Office through manual submission at the office address indicated below, on or before <i>March 11, 2024, 9:30 A.M.</i> Late bids shall not be accepted.</p> <p>xxx xxx xxx</p> <p>9. Bid opening shall be on <i>March 11, 2024, 10:00 A.M</i> at the given address below. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>	<p>AMENDMENT TO SECTION I. INVITATION TO BID</p> <p>xxx xxx xxx</p> <p>7. Bids must be duly received by the Procurement Management Office through manual submission at the office address indicated below, on or before March 15, 2024, 9:30 A.M. Late bids shall not be accepted.</p> <p>xxx xxx xxx</p> <p>9. Bid opening shall be on March 15, 2024, 10:00 A.M at the given address below. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>

	XXX XXX XXX	XXX XXX XXX
2	<p style="text-align: center;">PREVIOUS REMINDERS</p> <p style="text-align: center;">XXX XXX XXX</p> <p>BIDDING DOCUMENTS AND AVAILABILITY FEE</p> <ul style="list-style-type: none"> • Bidding Documents: <ul style="list-style-type: none"> ➤ February 19, 2024 to March 11, 2024 until 9:30 AM ➤ 8:00 am to 5:00 pm and upon payment of applicable fees for the Bidding Documents at the City Treasurer’s Office <p style="text-align: center;">XXX XXX XXX</p> <p>REMINDERS:</p> <ul style="list-style-type: none"> • The deadline for the submission of bid is on March 11, 2024 (Monday) at 9:30 AM at the BAC Secretariat’s Office, 4th Floor Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. The digital clock at the BAC Secretariat’s Office that is set to the Philippine Time (PhST) shall be used as reference in determining the time for the submission of bids, hence participating bidders are advised to synchronize their timepiece with the said digital clock. Late bids or those who submitted after 9:30 AM of March 11, 2024 (Monday) shall not be accepted. <p style="text-align: center;">XXX XXX XXX</p>	<p style="text-align: center;">AMENDMENT TO REMINDERS</p> <p style="text-align: center;">XXX XXX XXX</p> <p>BIDDING DOCUMENTS AND AVAILABILITY FEE</p> <ul style="list-style-type: none"> • Bidding Documents: <ul style="list-style-type: none"> ➤ February 19, 2024 to March 15, 2024 until 9:30 AM ➤ 8:00 am to 5:00 pm and upon payment of applicable fees for the Bidding Documents at the City Treasurer’s Office <p style="text-align: center;">XXX XXX XXX</p> <p>REMINDERS:</p> <ul style="list-style-type: none"> • The deadline for the submission of bid is on March 15, 2024 (Friday) at 9:30 AM at the BAC Secretariat’s Office, 4th Floor Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. The digital clock at the BAC Secretariat’s Office that is set to the Philippine Time (PhST) shall be used as reference in determining the time for the submission of bids, hence participating bidders are advised to synchronize their timepiece with the said digital clock. Late bids or those who submitted after 9:30 AM of March 15, 2024 (Friday) shall not be accepted. <p style="text-align: center;">XXX XXX XXX</p>


	<ul style="list-style-type: none"> Bid opening shall be on March 11, 2024 (Monday) at 10:00 AM at 7th Floor Meeting Room, Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. Bids will be opened in the presence of the bidders' representatives who choose to attend. <p style="text-align: center;">xxx xxx xxx</p>	<ul style="list-style-type: none"> Bid opening shall be on March 15, 2024 (Friday) at 10:00 AM at 7th Floor Meeting Room, Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. Bids will be opened in the presence of the bidders' representatives who choose to attend. <p style="text-align: center;">xxx xxx xxx</p>								
3	<p>PREVIOUS SECTION VII TECHNICAL SPECIFICATIONS</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 10%;">ITEM NO.</th> <th style="width: 90%;">SPECIFICATION</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">3</td> <td> Flexible Bronchoscope for Pedia with Processor, Monitor and Trolley/Stand. <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Monitor: Approx. 24 inches Insertion Tube: max 2.8 mm <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Bending (up/down): 210 Deg/130 Deg <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Lightened by copper, not fiber Processor: approx. 10.1-inches, resolution: 1280x800; dimension; approx. 250x182x45mm <div style="text-align: center;">xxx xxx xxx</div> </td> </tr> </tbody> </table>	ITEM NO.	SPECIFICATION	3	Flexible Bronchoscope for Pedia with Processor, Monitor and Trolley/Stand. <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Monitor: Approx. 24 inches Insertion Tube: max 2.8 mm <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Bending (up/down): 210 Deg/130 Deg <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Lightened by copper, not fiber Processor: approx. 10.1-inches, resolution: 1280x800; dimension; approx. 250x182x45mm <div style="text-align: center;">xxx xxx xxx</div>	<p>AMENDMENT TO SECTION VII TECHNICAL SPECIFICATIONS</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 10%;">ITEM NO.</th> <th style="width: 90%;">SPECIFICATION</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">3</td> <td> Flexible Bronchoscope for Pedia with Processor, Monitor / Pocket Monitor and Trolley/Stand. <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Monitor: Approx. 24 inches or Pocket Monitor at least 3.5 inches Insertion Tube: 3.5mm 2.8 mm <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Bending (up/down): 130-210 Deg/130 Deg <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Lightened by copper or LED Processor: 10-13 inches, Resolution: 1280-1980x800-1080mm <div style="text-align: center;">xxx xxx xxx</div> <p>Inclusion: Wi-fi module, power adapters and cables of monitor, rechargeable battery if applicable: cleaning materials for the scope.</p> </td> </tr> </tbody> </table>	ITEM NO.	SPECIFICATION	3	Flexible Bronchoscope for Pedia with Processor, Monitor / Pocket Monitor and Trolley/Stand. <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Monitor: Approx. 24 inches or Pocket Monitor at least 3.5 inches Insertion Tube: 3.5mm 2.8 mm <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Bending (up/down): 130-210 Deg/130 Deg <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Lightened by copper or LED Processor: 10-13 inches, Resolution: 1280-1980x800-1080mm <div style="text-align: center;">xxx xxx xxx</div> <p>Inclusion: Wi-fi module, power adapters and cables of monitor, rechargeable battery if applicable: cleaning materials for the scope.</p>
ITEM NO.	SPECIFICATION									
3	Flexible Bronchoscope for Pedia with Processor, Monitor and Trolley/Stand. <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Monitor: Approx. 24 inches Insertion Tube: max 2.8 mm <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Bending (up/down): 210 Deg/130 Deg <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Lightened by copper, not fiber Processor: approx. 10.1-inches, resolution: 1280x800; dimension; approx. 250x182x45mm <div style="text-align: center;">xxx xxx xxx</div>									
ITEM NO.	SPECIFICATION									
3	Flexible Bronchoscope for Pedia with Processor, Monitor / Pocket Monitor and Trolley/Stand. <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Monitor: Approx. 24 inches or Pocket Monitor at least 3.5 inches Insertion Tube: 3.5mm 2.8 mm <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Bending (up/down): 130-210 Deg/130 Deg <div style="text-align: center;">xxx xxx xxx</div> <ul style="list-style-type: none"> Lightened by copper or LED Processor: 10-13 inches, Resolution: 1280-1980x800-1080mm <div style="text-align: center;">xxx xxx xxx</div> <p>Inclusion: Wi-fi module, power adapters and cables of monitor, rechargeable battery if applicable: cleaning materials for the scope.</p>									

		XXX XXX XXX
4	<p>PREVIOUS TERMS OF REFERENCE</p> <p>XXX XXX XXX</p> <p>-At least three (3) Good Performance Certificate from end-user / institution to be submitted during post-qualification</p> <p>XXX XXX XXX</p>	<p>AMENDMENT TO TERMS OF REFERENCE</p> <p>XXX XXX XXX</p> <p>-At least two (2) Good Performance Certificate of any medical equipment from end-user / institution to be submitted during post-qualification</p> <p>XXX XXX XXX</p>

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as "Modified Bid") and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

Conforme:



 END-USER REPRESENTATIVE
 Signature Over Printed Name



ATTY. JOSEPHINE C. LATI-BAGAOISAN
 Chairperson