



BID BULLETIN NO. 01

Date: **19 April 2024**

ITB No.: **BAG-24-0401D**

Project Name: **Supply and Delivery of Various Medical Equipment – PCGH**

ABC: **PHP1,573,000.00**

To all prospective bidders:

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 01 April 2024. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.


Item	Previous Specification/ Clarification/Request to Consider	Amendment/Response to Clarification
1	<p>PREVIOUS SECTION I. INVITATION TO BID</p> <p>xxx xxx xxx</p> <p>7. Bids must be duly received by the Procurement Management Office through manual submission at the office address indicated below, on or before <i>22 April 2024, 9:30 A.M.</i> Late bids shall not be accepted.</p> <p>xxx xxx xxx</p> <p>9. Bid opening shall be on <i>22 April 2024, 10:00 A.M.</i> at the given address below. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>	<p>AMENDMENT TO SECTION I. INVITATION TO BID</p> <p>xxx xxx xxx</p> <p>7. Bids must be duly received by the Procurement Management Office through manual submission at the office address indicated below, on or before 26 April 2024, 9:30 A.M. Late bids shall not be accepted.</p> <p>xxx xxx xxx</p> <p>9. Bid opening shall be on 26 April 2024, 10:00 A.M. at the given address below. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>

	<p>7th Floor Meeting Room, Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. Bids will be opened in the presence of the bidders' representatives who choose to attend.</p> <p>xxx xxx xxx</p>	<p>7th Floor Meeting Room, Pasig City Hall, Caruncho Ave., San Nicolas Pasig City. Bids will be opened in the presence of the bidders' representatives who choose to attend.</p> <p>xxx xxx xxx</p>																
<p>3</p>	<p>PREVIOUS SECTION VII. TECHNICAL SPECIFICATIONS</p> <p>xxx xxx xxx</p> <table border="1"> <thead> <tr> <th data-bbox="308 633 411 696">Item</th> <th data-bbox="411 633 869 696">Specification</th> </tr> </thead> <tbody> <tr> <td data-bbox="308 696 411 1120">1</td> <td data-bbox="411 696 869 1120"> <p>INFUSION PUMP</p> <p>xxx xxx xxx</p> <p>4. The unit should have an accumulated infusion volume of 0.1- 9999.5 ml</p> <p>xxx xxx xxx</p> <p>7. The unit should have an infusion rate increment of 0.1 ml/hr</p> </td> </tr> <tr> <td data-bbox="308 1120 411 1153">xxxx</td> <td data-bbox="411 1120 869 1153"></td> </tr> <tr> <td data-bbox="308 1153 411 1668">3</td> <td data-bbox="411 1153 869 1668"> <p>12 LEAD ELECTROCARDIOGRAPHY WITH TROLLEY</p> <p>xxx xxx xxx</p> <p>c. Can print on standard paper or and A4 size thermal paper (215mm x 297mm)</p> <p>xxx xxx xxx</p> <p>d. Low pass filter (off, 40Hz, 100Hz, 150Hz)</p> <p>xxx xxx xxx</p> </td> </tr> </tbody> </table> <p>xxx xxx xxx</p>	Item	Specification	1	<p>INFUSION PUMP</p> <p>xxx xxx xxx</p> <p>4. The unit should have an accumulated infusion volume of 0.1- 9999.5 ml</p> <p>xxx xxx xxx</p> <p>7. The unit should have an infusion rate increment of 0.1 ml/hr</p>	xxxx		3	<p>12 LEAD ELECTROCARDIOGRAPHY WITH TROLLEY</p> <p>xxx xxx xxx</p> <p>c. Can print on standard paper or and A4 size thermal paper (215mm x 297mm)</p> <p>xxx xxx xxx</p> <p>d. Low pass filter (off, 40Hz, 100Hz, 150Hz)</p> <p>xxx xxx xxx</p>	<p>AMENDMENT TO SECTION VII. TECHNICAL SPECIFICATIONS</p> <p>xxx xxx xxx</p> <table border="1"> <thead> <tr> <th data-bbox="919 633 1023 696">Item</th> <th data-bbox="1023 633 1481 696">Specification</th> </tr> </thead> <tbody> <tr> <td data-bbox="919 696 1023 1120">1</td> <td data-bbox="1023 696 1481 1120"> <p>INFUSION PUMP</p> <p>xxx xxx xxx</p> <p>4. The unit should have an accumulated infusion volume of 0.1- 9999 ml</p> <p>xxx xxx xxx</p> <p>7. The unit should have an infusion rate increment of 0.01 ml/hr</p> </td> </tr> <tr> <td data-bbox="919 1120 1023 1153">xxxx</td> <td data-bbox="1023 1120 1481 1153"></td> </tr> <tr> <td data-bbox="919 1153 1023 1668">3</td> <td data-bbox="1023 1153 1481 1668"> <p>12 LEAD ELECTROCARDIOGRAPHY WITH TROLLEY</p> <p>xxx xxx xxx</p> <p>c. Can print on standard paper or and A4 size thermal paper (210mm x 297mm)</p> <p>xxx xxx xxx</p> <p>d. Muscle filter or any counterpart of low pass filter</p> <p>xxx xxx xxx</p> </td> </tr> </tbody> </table> <p>xxx xxx xxx</p>	Item	Specification	1	<p>INFUSION PUMP</p> <p>xxx xxx xxx</p> <p>4. The unit should have an accumulated infusion volume of 0.1- 9999 ml</p> <p>xxx xxx xxx</p> <p>7. The unit should have an infusion rate increment of 0.01 ml/hr</p>	xxxx		3	<p>12 LEAD ELECTROCARDIOGRAPHY WITH TROLLEY</p> <p>xxx xxx xxx</p> <p>c. Can print on standard paper or and A4 size thermal paper (210mm x 297mm)</p> <p>xxx xxx xxx</p> <p>d. Muscle filter or any counterpart of low pass filter</p> <p>xxx xxx xxx</p>
Item	Specification																	
1	<p>INFUSION PUMP</p> <p>xxx xxx xxx</p> <p>4. The unit should have an accumulated infusion volume of 0.1- 9999.5 ml</p> <p>xxx xxx xxx</p> <p>7. The unit should have an infusion rate increment of 0.1 ml/hr</p>																	
xxxx																		
3	<p>12 LEAD ELECTROCARDIOGRAPHY WITH TROLLEY</p> <p>xxx xxx xxx</p> <p>c. Can print on standard paper or and A4 size thermal paper (215mm x 297mm)</p> <p>xxx xxx xxx</p> <p>d. Low pass filter (off, 40Hz, 100Hz, 150Hz)</p> <p>xxx xxx xxx</p>																	
Item	Specification																	
1	<p>INFUSION PUMP</p> <p>xxx xxx xxx</p> <p>4. The unit should have an accumulated infusion volume of 0.1- 9999 ml</p> <p>xxx xxx xxx</p> <p>7. The unit should have an infusion rate increment of 0.01 ml/hr</p>																	
xxxx																		
3	<p>12 LEAD ELECTROCARDIOGRAPHY WITH TROLLEY</p> <p>xxx xxx xxx</p> <p>c. Can print on standard paper or and A4 size thermal paper (210mm x 297mm)</p> <p>xxx xxx xxx</p> <p>d. Muscle filter or any counterpart of low pass filter</p> <p>xxx xxx xxx</p>																	

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as "Modified Bid") and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

Conforme:


NORISSA G. SABARRÉ, MD FPCem
END-USER REPRESENTATIVE
Signature Over Printed Name


ATTY. JOSEPHINE C. LATI-BAGAOISAN
Chairperson