

**Bid Notice Abstract**

## Request for Quotation (RFQ)

Reference Number 9422225

Procuring Entity CITY OF PASIG

Title Supply and Delivery of Various Medical Equipment for the Pasig City General Hospital

Area of Delivery Metro Manila

Solicitation Number:	R1 100-22-10-6109	Status	Active
Trade Agreement:	Implementing Rules and Regulations	Associated Components	1
Procurement Mode:	Negotiated Procurement - Small Value Procurement (Sec. 53.9)	Bid Supplements	0
Classification:	Goods	Document Request List	0
Category:	Medical and Dental Equipment	Date Published	24/01/2023
Approved Budget for the Contract:	PHP 750,675.00	Last Updated / Time	24/01/2023 00:00 AM
Delivery Period:	60 Day/s	Closing Date / Time	27/01/2023 17:00 PM
Client Agency:			
Contact Person:	Rho Depaudhon BAC Secretariat Pasig City Hall, Caruncho Avenue,		

Barangay San Nicolas,  
Pasig City  
Metro Manila  
Philippines 1600  
63-2-86431111 Ext.1461

bidsandawards@pasigcity.gov.ph

Description

Items Quantity / Units

1 METZENBAUM SCISSORS, Technical Specifications:

\* Rotating, dismantling, insulated with connector pin for unipolar coagulation

\* with LUER-locking irrigation connector for cleaning

\* double action jaws, 15mm jaw curved, diameter 5mm, length 36cm

2 unit

2 NEEDLE HOLDER JAWS CURVED TO RIGHT, Technical Specification:

\* with Tungsten carbide insert ergonomic pistol handle with dis-engage able ratchet

\* Ratchet position right

\* Jaws curved to right, size 5mm, length 33cm

1 pcs

3 Heavy Duty Suction Machine, TECHNICAL SPECIFICATION:

" 2 pcs. 4L jar polysulfone (PSU) auto cleavable up to 135°C with optional rack for 3rd bottle

" Free Air Suction Flow: up to 91L/min.

" Maximum vacuum: 93 kPa- 933 mbar- 700 mmHg

" Dimensions: at least 900 x 500 x 470 mm ± 10mm range

" Power Supply: 200- 240 V (50-60 Hz)

" Power: 300 Watts

" Weight: 22- 25 kg

" Accessories: Bacteria filter 10pcs, Suction hose 6pcs, On/Off foot switch included

" Machine: High flow vacuum pump (up to 90L/min.), Double piston cylinder, oil free,

maintenance not required

" Material: made in polypropylene of high resistance, with 4 integrated wheels of Ø80mm,

two of them brake and on antistatic

" Noise level: 50 ±1,5 dB (AS) 1m

" Other requirements: Collection jars selector control knob with 2 positions (left or right).

On/Off switch and accurate system of vacuum regulation, with rotating control of smooth tact and easy cleaning.

Precision gauge of high vacuum range (0-100 kPa / 0-760 mmHg/ 0-14,7 Psi) with great visibility (Ø67 mm).

Lodging for the 3m electric wire.

2 unit

PRICE QUOTATION/S SHOULD BE PRINTED ON COMPANY'S OFFICIAL LETTERHEAD TOGETHER WITH THE FOLLOWING UPDATED DOCUMENTS;

-Mayor's/Business Permit

-PhilGEPS Registration Number

-Income/Business Tax Return

-Accomplished and notarized Omnibus Sworn Statement

(<https://www.gppb.gov.ph/assets/forms/Omnibus%20Sworn%20Statement> (Revised).docx)

-Proof of Authorization: Secretary's Certificate if corporation, or Special Power Of Attorney, if individual

NOTE:

TO BE SUBMITTED SEALED AND LABELED ON A LONG BROWN ENVELOPE, FOLLOWING THIS FORMAT AS FOLLOWS:

FOR:

ATTY. JOSEPHINE C. LATI-BAGAOISAN

BAC Chairperson

THRU:

ATTY. PONCE MIGUEL D. LOPEZ

Officer In Charge, Procurement Management Office

4th Floor, Pasig City Hall,

Caruncho Avenue,

Pasig City

DATE : \_\_\_\_\_

COMPANY'S NAME : \_\_\_\_\_

PhilGEPS REFERENCE NUMBER : \_\_\_\_\_

PROJECT TITLE : \_\_\_\_\_

Created by Rho Depaudhon

Date Created 23/01/2023

The PhilGEPS team is not responsible for any typographical errors or misinformation presented in the system. PhilGEPS only displays information provided for by its clients, and any queries regarding the postings should be directed to the contact person/s of the concerned party.