



LUNGSOD NG
PASIG
UMAAGOS ANG PAG-ASA

BIDS AND AWARDS COMMITTEE

BID BULLETIN NO. 01

Date: **10 May 2023**

ITB No.: **bac-23-0427b**

Project Name: **Supply and Delivery of One (1) Unit Portable Transport Ventilator for the Use of PCGH - Anesthesia Section**

ABC: **Php1,800,000.00**

To all prospective bidders:

This Bid Bulletin is issued to clarify, supplement, modify and/or revise the particular sections in the Bid and Contract Documents as stipulated in the Bidding Documents issued on 27 April 2023. The Bidders shall take note of the following items carefully and consider them in the preparation of their bid proposals, as they shall form part of the CONTRACT DOCUMENTS.

Item	Previous Specification/ Clarification/Request to Consider	Amendment/Response to Clarification												
1	PREVIOUS SECTION VII. TECHNICAL SPECIFICATION xxx xxx xxx <table border="1"><thead><tr><th>ITEM NO</th><th>SPECIFICATION</th></tr></thead><tbody><tr><td>3</td><td>Battery run time: not less than 5 hrs / Charging time: not greater than 5 hrs</td></tr><tr><td>8</td><td>Curves displayed: Volume and pressure flow curves</td></tr></tbody></table>	ITEM NO	SPECIFICATION	3	Battery run time: not less than 5 hrs / Charging time: not greater than 5 hrs	8	Curves displayed: Volume and pressure flow curves	AMENDMENT TO SECTION VII. TECHNICAL SPECIFICATION xxx xxx xxx <table border="1"><thead><tr><th>ITEM NO</th><th>SPECIFICATION</th></tr></thead><tbody><tr><td>3</td><td>Battery run time: If battery run time is less time 5 hours & Charging time greater than 5 hours, must provide a spare battery pack & separate charger</td></tr><tr><td>8</td><td>Curves displayed: Volume</td></tr></tbody></table>	ITEM NO	SPECIFICATION	3	Battery run time: If battery run time is less time 5 hours & Charging time greater than 5 hours, must provide a spare battery pack & separate charger	8	Curves displayed: Volume
ITEM NO	SPECIFICATION													
3	Battery run time: not less than 5 hrs / Charging time: not greater than 5 hrs													
8	Curves displayed: Volume and pressure flow curves													
ITEM NO	SPECIFICATION													
3	Battery run time: If battery run time is less time 5 hours & Charging time greater than 5 hours, must provide a spare battery pack & separate charger													
8	Curves displayed: Volume													

	xxx xxx xxx		or Pressure flow curves
			xxx xxx xxx
2	PREVIOUS ITEMS TO BE BID TOTAL: 1,860,000.00	AMENDMENT TO ITEMS TO BE BID TOTAL: 1,800,000.00	

Bidders who have already submitted bids are hereby informed that they are allowed to modify or withdraw their bids, if necessary, before the scheduled opening of bid envelopes.

For modifications in your original submitted bid, kindly submit new bidding documents (sealed and marked as "Modified Bid") and have these received at the Office of the Bids and Awards Secretariat. Bid modifications received after the deadline shall not be considered and shall be returned to the bidder unopened.

Conforme:

for: [Signature]
ROBERTO C. ABUEG MD, DPM, FPM

END-USER REPRESENTATIVE

Signature Over Printed Name

[Signature]
ATTY. JOSEPHINE C. LATI-BAGAOISAN
 Chairperson